



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Level 3 Inspection LLC & Smart Inspection Systems LLC
1239 SE Indian Street, Suite 107 & 108
Stuart, FL 34997

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

DIMENSIONAL MEASUREMENT

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AT-1791

Certificate Number



ANAB Approval

Certificate Valid Through: 06/13/2021
Version No. 006 Issued: 05/20/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Level 3 Inspection LLC & Smart Inspection Systems LLC

1239 SE Indian Street, Suite 107 & 108
Stuart, FL 34997

Scott McAfee Phone: (772) 427-6420
directors@level3inspection.com www.level3inspection.com

DIMENSIONAL MEASUREMENT

Valid to: June 13, 2021

Certificate Number: AT-1791

3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional -- Computer Aided Inspection (CAI) using High Accuracy 3D Scanning	FOV:		High Accuracy 3D Scanner
	50 mm	1.9 µm	Customer-defined specifications, blueprints, CAD model or requests
	100 mm	1.8 µm	
	200 mm	2.3 µm	
	400 mm	1.8 µm	
800 mm	1.7 µm		

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

1. FOV indicates cubic Field of View for scanner configuration.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-1791

Vice President