



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

**Level 3 Inspection LLC & Smart Inspection
Systems LLC**
1239 SE Indian Street, Suite 107 & 108
Stuart, FL 34997

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 13 June 2023
Certificate Number: AT-1791



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Level 3 Inspection LLC & Smart Inspection Systems LLC

1239 SE Indian Street, Suite 107 & 108
Stuart, FL 34997
Scott McAfee
772-427-6420

DIMENSIONAL MEASUREMENT

Valid to: **June 13, 2023**

Certificate Number: **AT-1791**

3 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional -- Computer Aided Inspection (CAI) using High Accuracy 3D Scanning	FOV:		High Accuracy 3D Scanner
	50 mm	1.9 µm	Customer-defined specifications, blueprints, CAD model or requests
	100 mm	1.8 µm	
	200 mm	2.3 µm	
	400 mm	1.8 µm	
800 mm	1.7 µm		

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. FOV indicates cubic Field of View for scanner configuration.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-1791.



R. Douglas Leonard Jr., VP, PILR SBU