



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**Level 3 Inspection, LLC**  
1239 SE Indian Street, Suite 108  
Stuart, FL 34997

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**DIMENSIONAL MEASUREMENT**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 13 June 2027

Certificate Number: AT-1791



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Level 3 Inspection, LLC**  
 1239 SE Indian Street, Suite 108  
 Stuart, FL 34997  
 Scott McAfee  
 772-214-4708

**DIMENSIONAL MEASUREMENT**

Valid to: **June 13, 2027**

Certificate Number: **AT-1791**

**3 Dimensional**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional -- Computer Aided Inspection (CAI) using High Accuracy 3D Scanning	FOV 50 mm (5 to 150) mm	2 µm	Measurement using High Accuracy 3D Scanner  Customer-defined specifications, blueprints, CAD model or requests
	FOV 100 mm (150 350) mm	2 µm	
	FOV 200 mm (150 350 X) mm	1.8 µm	
	FOV 400 mm (400 to 800) mm	2.8 µm	
	FOV 800 mm (400 to 800) mm	2.8 µm	
	FOV 800 mm (400 to 800) mm	2.8 µm	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. FOV indicates cubic Field of View for scanner configuration.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-1791.



Jason Stine, Vice President